

TwistAmp® exo Kit

Quick Guide

Part number: TAEXO02Guide

Revision F

TwistAmp[®] exo Quick Guide

Part number: TAEXO02Guide / Revision F

Please see **instruction** and **assay design manuals** at twistdx.co.uk for information regarding kit components and storage, assay design, detailed use and multiplexing.

Set-up (single-plex)

1. Prepare reaction mix in 1.5ml tube:

Primer A (10 μ M)	2.1 μ l
Primer B (10 μ M)	2.1 μ l
TwistAmp [®] exo probe (10 μ M)	0.6 μ l
Primer Free Rehydration buffer	29.5 μ l
Template and water to	13.2 μ l
(Total volume	47.5 μ l)
Vortex and spin briefly.	

2. Add reaction mix to a TwistAmp[®] exo reaction. Pipette to mix.

3. Add 2.5 μ l of 280mM Magnesium Acetate (MgOAc) (supplied) and mix well to start reaction.

Note: RPA reactions start as soon as MgOAc is added.

4. Place reactions in a fluorometer and start run: 39°C, 20 minutes. For low template copy number, remove strip after 4 minutes, vortex and spin briefly, replace in fluorometer.

Note: If tubes are opened after amplification there is a great risk of contamination of work surfaces with amplicon. Ensure that appropriate avoidance measures are taken.

TwistAmp® exo Quick Guide

Part number: TAEXO02Guide / Revision F

Kit Positive Control Set-up (single-plex)

1. Prepare reaction mix in 1.5ml tube:

Positive control primer/probe mix	8 μ l
Primer Free Rehydration buffer	29.5 μ l
Positive control DNA template	1 μ l
Water	9 μ l
(Total volume	47.5 μ l)

Vortex and spin briefly.

2. Add reaction mix to a TwistAmp® exo reaction. Pipette to mix.

3. Add 2.5 μ l of 280mM Magnesium Acetate (MgOAc) (supplied) and mix well to start reaction.

Note: RPA reactions start as soon as MgOAc is added.

4. Place reactions in a fluorometer and start run: 40°C, 20 minutes. Remove strip after 4 minutes, vortex and spin briefly, replace in fluorometer.

Note: If tubes are opened after amplification there is a great risk of contamination of work surfaces with amplicon. Ensure that appropriate avoidance measures are taken.

The exo positive control probe is labelled with a fluorescein (FAM) fluorophore, the excitation optimum is at 488nm and the emission maximum is at 520nm.

Notice to purchaser

Licence, Use Restrictions and Limitations of Liability

Definitions. As used in this section, “kit” means the items described in this manual (the “Manual”) and supplied by TwistDx to a purchaser (the “Recipient”). “Materials” means all biological and chemical materials supplied as part of the kit. “Information” means all written information supplied as part of the kit, information relating to the kit made available through TwistDx’s website, and any verbal or written information concerning the kit or its use provided by any employee or agent of TwistDx.

Limitations on Use and Distribution of the Material and Information

Recipient acknowledges and agrees that the Materials and Information are proprietary to TwistDx, may be covered by claims of patents or patent applications owned by TwistDx or its affiliates and are supplied subject to the following restrictions: The Materials and Information are non-exclusively licensed to Recipient solely for non-commercial internal research purposes and for applications other than the sequential determination of the identity and relative order of at least 200,000 total nucleotides in a single run on a sequencing apparatus. Any *in vitro* diagnostic use of the Materials and Information or any use for diagnosing or monitoring any medical condition in a human is expressly prohibited.

No Warranty; Limitation of Liability

Recipient understands and agrees that the materials are experimental in nature and that the Materials and Information are provided without any warranty as to results, merchantability, fitness for a particular purpose or non-infringement of any patent or other intellectual property right, and without any other representation, warranty or condition, express or implied. TwistDx shall not be liable in connection with the Materials, Information or any breach of this agreement under any contract, negligence, strict liability or other theory for (a) loss of revenues, loss of profits, or loss or inaccuracy of data, including test results, regardless of how such damages are characterised (b) for the cost (including procurement costs) of substitute goods, services or technology, or (c) for any special, indirect, incidental or consequential damages. In no event will TwistDx’s aggregate liability under this agreement exceed one hundred dollars (US\$100). Recipient understands that its use of the materials and/or the information in connection with its activities are entirely at its risk.